


Deliver a Bolus

for the t:slim X2 Insulin Pump



UNITS OF MEASURE MMOL/L


The instructions below are provided as a reference tool for caregivers who are already familiar with the use of an insulin pump and with insulin therapy in general. Not all screens are shown. For more detailed information on the operation of your Tandem insulin pump, please refer to its user guide.



1.

Tap **0 grams** to enter the carbs for your bolus.


NOTE: If this area reads “units,” the carb feature is turned off in the active profile.



2.

Enter desired value. Be sure “grams” is displayed above keypad for food boluses.


Tap to continue.



3.


Tap **Add BG** to enter your blood glucose (BG).

NOTE: If you have a CGM session active, and if there is both a CGM value and a CGM trend arrow available on the CGM Home Screen, your glucose value is autopopulated in the GLUCOSE field.




4.

Enter desired value. Be sure “mmol/L” is displayed above keypad when entering BG values.




5.

If a BG is entered that is below the target, but above or 3.9 mmol/L, you will be offered the option to reduce the bolus amount. To accept that reduction tap ; otherwise, tap .



6.


Tap to continue. Tap the calculated units value to manually adjust recommended dose.



7.

Verify the dose and tap to confirm.


NOTE: Calculations above are based on preset insulin-to-carb ratios and correction factors, which may be set in Personal Profiles.



8.

Tap to deliver the food bolus immediately.

The BOLUS INITIATED screen will appear to confirm delivery has started.



9.

To cancel the undelivered portion of the bolus, tap next to BOLUS on the Home Screen, then tap to confirm cancelled bolus.



TANDEM™
DIABETES CARE

(833) 509-3598
tandemdiabetes.ca



t:simulator App

A demo as simple as our pumps!



FOR USE IN CANADA ONLY

